

Surrogate Summary

SDG No.: **Q1235**

Client: **RU2 Engineering, LLC**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093980.D	PIBLK-PL093980.D	Decachlorobiphenyl	1	20	21.6	108		43	140
		Tetrachloro-m-xylene	1	20	22.2	111		77	126
		Decachlorobiphenyl	2	20	20.2	101		43	140
		Tetrachloro-m-xylene	2	20	21.8	109		77	126
PB166427BL	PB166427BL	Decachlorobiphenyl	1	20	21.1	106		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
		Decachlorobiphenyl	2	20	14.5	73		43	140
		Tetrachloro-m-xylene	2	20	17.5	87		77	126
PB166427BS	PB166427BS	Decachlorobiphenyl	1	20	23.4	117		43	140
		Tetrachloro-m-xylene	1	20	23.7	119		77	126
		Decachlorobiphenyl	2	20	20.9	105		43	140
		Tetrachloro-m-xylene	2	20	23.0	115		77	126
PB166356TB	PB166356TB	Decachlorobiphenyl	1	20	20.1	101		43	140
		Tetrachloro-m-xylene	1	20	21.0	105		77	126
		Decachlorobiphenyl	2	20	18.6	93		43	140
		Tetrachloro-m-xylene	2	20	19.5	98		77	126
Q1215-04MS	JPP-29.1-012825MS	Decachlorobiphenyl	1	20	22.9	115		43	140
		Tetrachloro-m-xylene	1	20	20.0	100		77	126
		Decachlorobiphenyl	2	20	20.6	103		43	140
		Tetrachloro-m-xylene	2	20	19.2	96		77	126
Q1215-04MSD	JPP-29.1-012825MSD	Decachlorobiphenyl	1	20	22.9	114		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	21.7	108		43	140
		Tetrachloro-m-xylene	2	20	20.1	100		77	126
I.BLK-PL093999.D	PIBLK-PL093999.D	Decachlorobiphenyl	1	20	28.1	140		43	140
		Tetrachloro-m-xylene	1	20	21.1	106		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	102		77	126
Q1235-04	JPP-51.2-012925	Decachlorobiphenyl	1	20	26.6	133		43	140
		Tetrachloro-m-xylene	1	20	25.5	127	*	77	126
		Decachlorobiphenyl	2	20	27.6	138		43	140
		Tetrachloro-m-xylene	2	20	24.9	124		77	126
Q1235-08	JPP-16.1-012925	Decachlorobiphenyl	1	20	25.1	126		43	140
		Tetrachloro-m-xylene	1	20	25.2	126		77	126
		Decachlorobiphenyl	2	20	26.4	132		43	140
		Tetrachloro-m-xylene	2	20	24.2	121		77	126
I.BLK-PL094020.D	PIBLK-PL094020.D	Decachlorobiphenyl	1	20	22.8	114		43	140

Surrogate Summary

SDG No.: Q1235

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094020.D	PIBLK-PL094020.D	Tetrachloro-m-xylene	1	20	21.6	108		77	126
		Decachlorobiphenyl	2	20	24.3	121		43	140
		Tetrachloro-m-xylene	2	20	21.1	106		77	126