

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1235 SAS No.: Q1235 SDG NO.: Q1235

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/05/2025

Lab File ID: BF141435.D Time Analyzed: 10:02

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	78610	6.792	301871	8.08	162758	9.83
UPPER LIMIT	157220	7.292	603742	8.575	325516	10.328
LOWER LIMIT	39305	6.292	150936	7.575	81379	9.328
EPA SAMPLE NO.						
01 PB166418BL	76450	6.79	302478	8.07	163001	9.83
02 PB166418BS	69377	6.79	275008	8.08	148846	9.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.