

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM

SAS No.: Q1235

SDG NO.: Q1235

Lab File ID: BF141372.D

DFTPP Injection Date: 02/03/2025

Instrument ID: BNA_F

DFTPP Injection Time: 16:43

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	44.1
68	Less than 2.0% of mass 69	0.8 (2) 1
69	Mass 69 relative abundance	37.8
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	50.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	26.7
365	Greater than 1% of mass 198	3.2
441	Present, but less than mass 443	12.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.7 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF141373.D	02/03/2025	17:14
JPP-51.2-012925	Q1235-04	BF141378.D	02/03/2025	19:31
JPP-16.1-012925	Q1235-08	BF141379.D	02/03/2025	19:57