

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1235

Client: RU2 Engineering, LLC

Analytical Method: SW8260D

Datafile : VX044806.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0203WBSD01	Vinyl chloride	20	19.2	ug/L	96	5		65	117	20
	1,1-Dichloroethene	20	20.4	ug/L	102	19		74	110	20
	2-Butanone	100	110	ug/L	110	0		65	122	20
	Carbon Tetrachloride	20	17.7	ug/L	89	1		77	113	20
	Chloroform	20	19.5	ug/L	98	1		79	113	20
	Benzene	20	20.8	ug/L	104	0		82	109	20
	1,2-Dichloroethane	20	18.3	ug/L	92	2		80	115	20
	Trichloroethene	20	20.0	ug/L	100	3		77	113	20
	Tetrachloroethene	20	18.7	ug/L	94	1		67	123	20
	Chlorobenzene	20	20.1	ug/L	101	1		82	109	20