

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1235 SAS No.: Q1235 SDG NO.: Q1235
 Lab File ID: VY021046.D Date Analyzed: 02/04/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:47
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	102470	13.352				
UPPER LIMIT	204940	13.852				
LOWER LIMIT	51235	12.852				
EPA SAMPLE NO.						
JPP-51.2-012925	111552	13.35				
JPP-16.1-012925	75300	13.35				
VY0204SBL01	114535	13.35				
VY0204SBS01	108974	13.35				
VY0204SBSD01	101483	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.