

Continuing Calibration Report for SequenceID : FD013125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049035.D

Aliphatic C9-C12	10200366.000	60.000	3.006	6.243	170006.100	162866.213	-4.384
Aliphatic C12-C16	7108969.000	40.000	6.244	9.612	177724.225	168872.221	-5.242
Aliphatic C16-C21	10782546.000	60.000	9.613	12.953	179709.100	172902.817	-3.936
Aliphatic C21-C28	13821355.000	80.000	12.954	16.592	172766.938	168565.180	-2.493
Aliphatic C28-C40	16117521.000	120.000	16.593	21.288	134312.675	131483.064	-2.152
Aliphatic EPH	58030757.000	360.000	3.006	21.288	161196.547	156011.702	-3.323