

Continuing Calibration Report for SequenceID : FD020325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049038.D

Aliphatic C9-C12	9798435.000	60.000	3.004	6.241	163307.250	162866.213	-0.271
Aliphatic C12-C16	6867752.000	40.000	6.242	9.610	171693.800	168872.221	-1.671
Aliphatic C16-C21	10471077.000	60.000	9.611	12.951	174517.950	172902.817	-0.934
Aliphatic C21-C28	13437689.000	80.000	12.952	16.589	167971.113	168565.180	0.352
Aliphatic C28-C40	15654279.000	120.000	16.590	21.285	130452.325	131483.064	0.784
Aliphatic EPH	56229232.000	360.000	3.004	21.285	156192.311	156011.702	-0.116