

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: SCIA01
ProjectID: 1063 Lafayette Ave, Hawthorne, NJ
Lab Code: CHEM Case No.: Q1236 SAS No.: Q1236 SDG No.: Q1236
DataFile: FE052201.D Analyst Name: YP\AJ Analyst Date: 02-01-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	78732712	92627	98925	6.366