

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: SCIA01

Lab Code: CHEM Case No.: Q1236 SAS No.: Q1236 SDG No.: Q1236

Instrument ID: MSVOA_D Calibration Date/Time: 02/04/2025 11:55

Lab File ID: VD080233.D Init. Calib. Date(s): 02/03/2025 02/03/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:27 12:42

GC Column: RTX-VMS ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Dichlorodifluoromethane	50.000	43.61	12.79	20
Bromomethane	50.000	48.21	3.573997	20