

Instrument ID: MSVOA\_D

**Daily Analysis Runlog For Sequence/QC Batch ID # VD020325**

|              |                     |                   |                                   |
|--------------|---------------------|-------------------|-----------------------------------|
| Review By    | Mahesh Dadoda       | Review On         | 2/4/2025 3:46:42 PM               |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 2/4/2025 3:49:18 PM               |
| SubDirectory | VD020325            | HP Acquire Method | HP Processing Method 82D020325s.m |

| STD. NAME                | STD REF.#                                             |
|--------------------------|-------------------------------------------------------|
| Tune/Reschk              | VP132840                                              |
| Initial Calibration Stds | VP132838,VP132842,VP132844,VP132846,VP132848,VP132850 |
| CCC                      | VP132854                                              |
| Internal Standard/PEM    | VP131783                                              |
| ICV/I.BLK                | VP132852                                              |
| Surrogate Standard       |                                                       |
| MS/MSD Standard          |                                                       |
| LCS Standard             |                                                       |

| Sr# | SampleID     | ClientID     | Data File Name | Date-Time         | Comment                                        | Operator | Status |
|-----|--------------|--------------|----------------|-------------------|------------------------------------------------|----------|--------|
| 1   | BFB          | BFB          | VD080218.D     | 03 Feb 2025 09:37 |                                                | RP/MD    | Ok     |
| 2   | VSTDICC005   | VSTDICC005   | VD080219.D     | 03 Feb 2025 10:27 | %D Failed for com.#05 in 5 ppb                 | RP/MD    | Ok,M   |
| 3   | VSTDICC010   | VSTDICC010   | VD080220.D     | 03 Feb 2025 10:54 | Com.#02 on LR And Com.#05,10,95 is on QR       | RP/MD    | Ok,M   |
| 4   | VSTDICC020   | VSTDICC020   | VD080221.D     | 03 Feb 2025 11:21 |                                                | RP/MD    | Ok,M   |
| 5   | VSTDICCC050  | VSTDICCC050  | VD080222.D     | 03 Feb 2025 11:48 |                                                | RP/MD    | Ok,M   |
| 6   | VSTDICC100   | VSTDICC100   | VD080223.D     | 03 Feb 2025 12:15 |                                                | RP/MD    | Ok,M   |
| 7   | VSTDICC150   | VSTDICC150   | VD080224.D     | 03 Feb 2025 12:42 |                                                | RP/MD    | Ok,M   |
| 8   | VIBLK        | VIBLK        | VD080225.D     | 03 Feb 2025 13:10 |                                                | RP/MD    | Ok     |
| 9   | VSTDICV050   | ICVVD020325  | VD080226.D     | 03 Feb 2025 14:15 | ICV failed for com.#58                         | RP/MD    | Ok,M   |
| 10  | VD0203SBL01  | VD0203SBL01  | VD080227.D     | 03 Feb 2025 15:12 |                                                | RP/MD    | Ok     |
| 11  | VD0203SBS01  | VD0203SBS01  | VD080228.D     | 03 Feb 2025 16:05 |                                                | RP/MD    | Ok,M   |
| 12  | VD0203SBSD01 | VD0203SBSD01 | VD080229.D     | 03 Feb 2025 16:33 |                                                | RP/MD    | Ok,M   |
| 13  | Q1271-01     | RBR200030    | VD080230.D     | 03 Feb 2025 17:57 | VIAL- A Internal standard fail; Surrogate fail | RP/MD    | ReRun  |

M : Manual Integration

Instrument ID: MSVOA\_D

**Daily Analysis Runlog For Sequence/QC Batch ID # VD020425**

|              |                     |                   |                                   |
|--------------|---------------------|-------------------|-----------------------------------|
| Review By    | Maresh Dadoda       | Review On         | 2/6/2025 11:05:33 AM              |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 2/6/2025 11:07:33 AM              |
| SubDirectory | VD020425            | HP Acquire Method | HP Processing Method 82y020325s.m |

| STD. NAME                               | STD REF.#         |
|-----------------------------------------|-------------------|
| Tune/Reschk<br>Initial Calibration Stds | VP132862          |
| CCC                                     | VP132863,VP132864 |
| Internal Standard/PEM                   | VP131783          |
| ICV/I.BLK                               |                   |
| Surrogate Standard                      |                   |
| MS/MSD Standard                         |                   |
| LCS Standard                            |                   |

| Sr# | SampleID     | ClientID       | Data File Name | Date-Time            | Comment                       | Operator | Status |
|-----|--------------|----------------|----------------|----------------------|-------------------------------|----------|--------|
| 1   | BFB          | BFB            | VD080232.D     | 04 Feb 2025 09:25    |                               | RP/MD    | Ok     |
| 2   | VSTDCCC050   | VSTDCCC050     | VD080233.D     | 04 Feb 2025 11:55    |                               | RP/MD    | Ok,M   |
| 3   | VD0204SBL01  | VD0204SBL01    | VD080234.D     | 04 Feb 2025 12:31    |                               | RP/MD    | Ok     |
| 4   | VD0204SBS01  | VD0204SBS01    | VD080235.D     | 04 Feb 2025 01:02 pm |                               | RP/MD    | Ok,M   |
| 5   | VD0204SBSD01 | VD0204SBSD01   | VD080236.D     | 04 Feb 2025 01:30 pm |                               | RP/MD    | Ok,M   |
| 6   | Q1254-01     | OK-02-01312025 | VD080237.D     | 04 Feb 2025 15:25    | Internal Standard Fail vial-A | RP/MD    | Ok     |
| 7   | Q1236-02     | VOC            | VD080238.D     | 04 Feb 2025 15:53    | Internal Standard Fail vial-B | RP/MD    | Ok     |
| 8   | Q1277-01     | RT 3249-VOC    | VD080239.D     | 04 Feb 2025 16:33    | Not purge vial-A              | RP/MD    | Not Ok |

M : Manual Integration