

Surrogate Summary

SDG No.: Q1236

Client: Sciacca General Contractors, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1236-02	VOC	1,2-Dichloroethane-d4	50	71.6	143 *	70 (50)	130 (163)
		Dibromofluoromethane	50	61.0	122	70 (54)	130 (147)
		Toluene-d8	50	39.6	79	70 (58)	130 (134)
		4-Bromofluorobenzene	50	39.5	79	70 (29)	130 (146)
VD0204SBL01	VD0204SBL01	1,2-Dichloroethane-d4	50	53.8	108	70 (50)	130 (163)
		Dibromofluoromethane	50	51.4	103	70 (54)	130 (147)
		Toluene-d8	50	48.5	97	70 (58)	130 (134)
		4-Bromofluorobenzene	50	44.2	88	70 (29)	130 (146)
VD0204SBS01	VD0204SBS01	1,2-Dichloroethane-d4	50	52.8	106	70 (50)	130 (163)
		Dibromofluoromethane	50	55.5	111	70 (54)	130 (147)
		Toluene-d8	50	56.8	114	70 (58)	130 (134)
		4-Bromofluorobenzene	50	58.7	117	70 (29)	130 (146)
VD0204SBSD01	VD0204SBSD01	1,2-Dichloroethane-d4	50	56.2	112	70 (50)	130 (163)
		Dibromofluoromethane	50	54.7	109	70 (54)	130 (147)
		Toluene-d8	50	56.4	113	70 (58)	130 (134)
		4-Bromofluorobenzene	50	58.2	116	70 (29)	130 (146)

