

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCIA01
 Lab Code: CHEM Case No.: Q1236 SAS No.: Q1236 SDG NO.: Q1236
 Lab File ID: VD080233.D Date Analyzed: 02/04/2025
 Instrument ID: MSVOA_D Time Analyzed: 11:55
 GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	144407	7.88	231304	8.78	218761	11.58
UPPER LIMIT	288814	8.375	462608	9.281	437522	12.081
LOWER LIMIT	72203.5	7.375	115652	8.281	109381	11.081
EPA SAMPLE NO.						
VOC	7896 *	7.88	14310 *	8.78	13671 *	11.58
VD0204SBL01	108713	7.88	188910	8.78	172019	11.58
VD0204SBS01	120560	7.88	198480	8.78	187950	11.58
VD0204SBSD01	119247	7.88	199616	8.78	189350	11.58

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.