

Continuing Calibration Report for SequenceID : FC013125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068171.D

Aliphatic C9-C12	7772727.000	60.000	3.128	6.394	129545.450	138488.174	6.457
Aliphatic C12-C16	5123126.000	40.000	6.395	9.777	128078.150	137713.071	6.996
Aliphatic C16-C21	7519004.000	60.000	9.778	13.129	125316.733	135019.289	7.186
Aliphatic C21-C28	9844182.000	80.000	13.130	16.780	123052.275	127207.526	3.267
Aliphatic C28-C40	13881415.000	120.000	16.781	21.588	115678.458	103638.860	-11.617
Aliphatic EPH	44140454.000	360.000	3.128	21.588	122612.372	123700.655	0.880