

Continuing Calibration Report for SequenceID : FF020125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015427.D

Aliphatic C9-C12	7864497.000	60.000	3.214	6.833	131074.950	133293.048	1.664
Aliphatic C12-C16	5485868.000	40.000	6.834	10.268	137146.700	138403.375	0.908
Aliphatic C16-C21	7976523.000	60.000	10.269	13.636	132942.050	144457.893	7.972
Aliphatic C21-C28	9708387.000	80.000	13.637	17.303	121354.838	142415.321	14.788
Aliphatic C28-C40	12710386.000	120.000	17.304	22.207	105919.883	104686.549	-1.178
Aliphatic EPH	43745661.000	360.000	3.214	22.207	121515.725	128213.342	5.224