

Continuing Calibration Report for SequenceID : FF020325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015430.D

Aliphatic C9-C12	8714399.000	60.000	3.213	6.835	145239.983	133293.048	-8.963
Aliphatic C12-C16	6093184.000	40.000	6.836	10.270	152329.600	138403.375	-10.062
Aliphatic C16-C21	8984906.000	60.000	10.271	13.640	149748.433	144457.893	-3.662
Aliphatic C21-C28	10909299.000	80.000	13.641	17.307	136366.238	142415.321	4.247
Aliphatic C28-C40	12159714.000	120.000	17.308	22.219	101330.950	104686.549	3.205
Aliphatic EPH	46861502.000	360.000	3.213	22.219	130170.839	128213.342	-1.527