

Continuing Calibration Report for SequenceID : FF020325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015435.D

Aliphatic C9-C12	8006762.000	60.000	3.213	6.835	133446.033	133293.048	-0.115
Aliphatic C12-C16	5498573.000	40.000	6.836	10.270	137464.325	138403.375	0.678
Aliphatic C16-C21	8038176.000	60.000	10.271	13.640	133969.600	144457.893	7.260
Aliphatic C21-C28	9540420.000	80.000	13.641	17.307	119255.250	142415.321	16.262
Aliphatic C28-C40	11027048.000	120.000	17.308	22.219	91892.067	104686.549	12.222
Aliphatic EPH	42110979.000	360.000	3.213	22.219	116974.942	128213.342	8.765