

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1237 SAS No.: Q1237 SDG NO.: Q1237

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/13/2025

Lab File ID: BF141603.D Time Analyzed: 10:07

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	101199	6.787	394924	8.07	212437	9.83
UPPER LIMIT	202398	7.287	789848	8.569	424874	10.328
LOWER LIMIT	50599.5	6.287	197462	7.569	106219	9.328
EPA SAMPLE NO.						
01 HL2PX3	97501	6.79	397436	8.06	223685	9.82
02 WC-13MS	96324	6.79	378872	8.07	214420	9.82
03 WC-13MSD	100491	6.79	409269	8.07	227525	9.82

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.