

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: GENV01

Lab Code: CHEM

SAS No.: Q1237

SDG NO.: Q1237

Lab File ID: BF141550.D

DFTPP Injection Date: 02/11/2025

Instrument ID: BNA\_F

DFTPP Injection Time: 10:02

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	41.7
68	Less than 2.0% of mass 69	0.7 ( 1.9 ) 1
69	Mass 69 relative abundance	36.4
70	Less than 2.0% of mass 69	0.2 ( 0.7 ) 1
127	10.0 - 80.0% of mass 198	48.6
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	28.2
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	13.1
442	Greater than 50% of mass 198	81.9
443	15.0 - 24.0% of mass 442	15.6 ( 19.1 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF141551.D	02/11/2025	10:28
PB166640BL	PB166640BL	BF141552.D	02/11/2025	10:55
PB166640BS	PB166640BS	BF141553.D	02/11/2025	11:21