

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q1238
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q1238 **SAS No.:** Q1238
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	981	1000	98	90 - 110	P	02/06/2025	12:59	Lb134604

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q1238
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q1238 **SAS No.:** Q1238
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.1	12.0	92	80 - 120	P	02/06/2025	13:03	Lb134604

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q1238
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q1238 **SAS No.:** Q1238
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4870	5000	98	90 - 110	P	02/06/2025	13:25	Lb134604
CCV02	Lead	4850	5000	97	90 - 110	P	02/06/2025	13:59	Lb134604
CCV03	Lead	4740	5000	95	90 - 110	P	02/06/2025	14:51	Lb134604
CCV04	Lead	4920	5000	98	90 - 110	P	02/06/2025	15:09	Lb134604

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q1238
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q1238 **SAS No.:** Q1238
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	999	1000	100	95 - 105	P	02/07/2025	12:11	LB134636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q1238
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q1238 **SAS No.:** Q1238
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.2	12.0	94	80 - 120	P	02/07/2025	12:27	LB134636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc. **SDG No.:** Q1238

Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q1238 **SAS No.:** Q1238

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4990	5000	100	90 - 110	P	02/07/2025	12:57	LB134636
CCV02	Lead	4970	5000	99	90 - 110	P	02/07/2025	13:49	LB134636
CCV03	Lead	4920	5000	98	90 - 110	P	02/07/2025	14:41	LB134636
CCV04	Lead	4890	5000	98	90 - 110	P	02/07/2025	15:32	LB134636
CCV05	Lead	4990	5000	100	90 - 110	P	02/07/2025	16:24	LB134636
CCV06	Lead	4910	5000	98	90 - 110	P	02/07/2025	17:16	LB134636
CCV07	Lead	4980	5000	100	90 - 110	P	02/07/2025	18:00	LB134636
CCV08	Lead	5020	5000	100	90 - 110	P	02/07/2025	18:23	LB134636
