



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Total Quality Environmental Inc.

SDG No.: Q1238

Contract: TOTA02

Lab Code: CHEM

Case No.: Q1238

SAS No.: Q1238

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Lead	11.4	12.0	95	40 - 160	P	02/06/2025	13:12	Lb134604
CRI01	Lead	10.9	12.0	91	40 - 160	P	02/07/2025	12:35	LB134636