

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Total Quality Environmental Inc. **SDG No.:** Q1238
Contract: TOTA02 **Lab Code:** CHEM **Case No.:** Q1238 **SAS No.:** Q1238
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	6.57			-12	12	02/06/2025	13:16	Lb134604
ICSAB01	Lead	56.8	49.0	116	37	61	02/06/2025	13:21	Lb134604
ICSA01	Lead	8.02			-12	12	02/07/2025	12:40	LB134636
ICSAB01	Lead	56.3	49.0	115	37	61	02/07/2025	12:44	LB134636