

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Total Quality Environmental Inc.</u>	level: <u>low</u>	sdg no.: <u>Q1238</u>
contract: <u>TOTA02</u>	lab code: <u>CHEM</u>	case no.: <u>Q1238</u> sas no.: <u>Q1238</u>
matrix: <u>Water</u>	sample id: <u>Q1262-05</u>	client id: <u>3762MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q1262-05MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	4580		60.0	U	5000	92		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Total Quality Environmental Inc.</u>	level: <u>low</u>	sdg no.: <u>Q1238</u>
contract: <u>TOTA02</u>	lab code: <u>CHEM</u>	case no.: <u>Q1238</u> sas no.: <u>Q1238</u>
matrix: <u>Water</u>	sample id: <u>Q1262-05</u>	client id: <u>3762MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q1262-05MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	4560		60.0	U	5000	91		P