

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1241

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166423TB		WATER		Batch Number:	PB166489		Prep Date:	02/03/2025	
	Mercury	2.00	<2.00	U	2.00	CV	02/04/2025	12:34	LB134551
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166489BL		WATER		Batch Number:	PB166489		Prep Date:	02/03/2025	
	Mercury	0.20	<0.20	U	0.20	CV	02/04/2025	11:45	LB134551

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1241

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166423TB		WATER		Batch Number:	PB166488		Prep Date:	02/03/2025	
	Arsenic	100	<100	U	100	P	02/07/2025	13:22	LB134636
	Barium	500	<500	U	500	P	02/07/2025	13:22	LB134636
	Cadmium	30.0	<30.0	U	30.0	P	02/07/2025	13:22	LB134636
	Chromium	50.0	<50.0	U	50.0	P	02/07/2025	13:22	LB134636
	Lead	60.0	<60.0	U	60.0	P	02/07/2025	13:22	LB134636
	Selenium	100	<100	U	100	P	02/07/2025	13:22	LB134636
	Silver	50.0	<50.0	U	50.0	P	02/07/2025	13:22	LB134636
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166488BL		WATER		Batch Number:	PB166488		Prep Date:	02/03/2025	
	Arsenic	100	<100	U	100	P	02/07/2025	14:15	LB134636
	Barium	500	<500	U	500	P	02/07/2025	14:15	LB134636
	Cadmium	30.0	<30.0	U	30.0	P	02/07/2025	14:15	LB134636
	Chromium	50.0	<50.0	U	50.0	P	02/07/2025	14:15	LB134636
	Lead	60.0	<60.0	U	60.0	P	02/07/2025	14:15	LB134636
	Selenium	100	<100	U	100	P	02/07/2025	14:15	LB134636
	Silver	50.0	<50.0	U	50.0	P	02/07/2025	14:15	LB134636