

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1241

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1241

SAS No.: Q1241

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166488BS							
Arsenic	ug/L	4000	4010		100	80 - 120	P
Barium	ug/L	1000	895		90	80 - 120	P
Cadmium	ug/L	1000	973		97	80 - 120	P
Chromium	ug/L	2000	2040		102	80 - 120	P
Lead	ug/L	5000	4820		96	80 - 120	P
Selenium	ug/L	10000	10200		102	80 - 120	P
Silver	ug/L	380	376		99	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: RU2 Engineering, LLC

SDG No.: Q1241

Contract: RUTW01

Lab Code: CHEM

Case No.: Q1241

SAS No.: Q1241

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166489BS Mercury	ug/L	4.0	4.23		106	80 - 120	CV