

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/03/2025 02/03/2025

Client Sample No.: CCAL06 Date Analyzed: 02/04/2025

Lab Sample No.: AR1660CCC500 Data File : PO109422.D Time Analyzed: 13:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.791	4.693	4.893	458.750	500.000	-8.3
Aroclor-1016-2	4.811	4.713	4.913	458.110	500.000	-8.4
Aroclor-1016-3	4.867	4.769	4.969	448.590	500.000	-10.3
Aroclor-1016-4	4.988	4.890	5.090	439.950	500.000	-12.0
Aroclor-1016-5	5.245	5.148	5.348	437.600	500.000	-12.5
Aroclor-1260-1	6.287	6.190	6.390	343.690	500.000	-31.3
Aroclor-1260-2	6.476	6.378	6.578	344.860	500.000	-31.0
Aroclor-1260-3	6.845	6.747	6.947	340.860	500.000	-31.8
Aroclor-1260-4	7.105	7.007	7.207	323.390	500.000	-35.3
Aroclor-1260-5	7.346	7.249	7.449	316.680	500.000	-36.7
Decachlorobiphenyl	8.755	8.658	8.858	35.320	50.000	-29.4
Tetrachloro-m-xylene	3.698	3.599	3.799	48.980	50.000	-2.0

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/03/2025 02/03/2025

Client Sample No.: CCAL06 Date Analyzed: 02/04/2025

Lab Sample No.: AR1660CCC500 Data File : PO109422.D Time Analyzed: 13:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.778	4.679	4.879	474.730	500.000	-5.1
Aroclor-1016-2	4.797	4.698	4.898	478.930	500.000	-4.2
Aroclor-1016-3	4.972	4.874	5.074	466.200	500.000	-6.8
Aroclor-1016-4	5.014	4.915	5.115	453.590	500.000	-9.3
Aroclor-1016-5	5.228	5.129	5.329	466.150	500.000	-6.8
Aroclor-1260-1	6.261	6.162	6.362	383.820	500.000	-23.2
Aroclor-1260-2	6.448	6.349	6.549	385.730	500.000	-22.9
Aroclor-1260-3	6.601	6.503	6.703	356.180	500.000	-28.8
Aroclor-1260-4	7.073	6.974	7.174	353.890	500.000	-29.2
Aroclor-1260-5	7.314	7.214	7.414	353.190	500.000	-29.4
Decachlorobiphenyl	8.707	8.607	8.807	38.440	50.000	-23.1
Tetrachloro-m-xylene	3.694	3.595	3.795	52.660	50.000	5.3