

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020325\
 Data File : P0109394.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2025 01:35
 Operator : YP/AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 02:50:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0020325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 04 02:49:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.699	3.695	577.5E6	400.0E6	73.025	72.858
2) SA Decachlor...	8.760	8.710	761.8E6	387.2E6	72.918	71.760
Target Compounds						
41) L9 AR-1268-1	7.635	7.601	782.3E6	436.2E6	733.993	730.415
42) L9 AR-1268-2	7.700	7.665	720.7E6	399.2E6	733.936	728.243
43) L9 AR-1268-3	7.909	7.873	591.7E6	318.5E6	733.714	728.436
44) L9 AR-1268-4	8.196	8.157	237.5E6	125.0E6	710.463	719.717
45) L9 AR-1268-5	8.495	8.452	1671.4E6	854.6E6	739.041	735.796



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020325\
Data File : PO109394.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2025 01:35
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 02:50:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO020325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 04 02:49:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.699 min
Delta R.T.: 0.000 min
Response: 577531401
Conc: 73.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.695 min
Delta R.T.: 0.000 min
Response: 400015610
Conc: 72.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.760 min
Delta R.T.: 0.001 min
Response: 761784514
Conc: 72.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.710 min
Delta R.T.: 0.001 min
Response: 387210924
Conc: 71.76 ng/ml



#41 AR-1268-1


R.T.: 7.635 min
Delta R.T.: 0.000 min
Response: 782288015
Conc: 733.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#41 AR-1268-1

R.T.: 7.601 min
Delta R.T.: 0.000 min
Response: 436226211
Conc: 730.42 ng/ml




#42 AR-1268-2

R.T.: 7.700 min
Delta R.T.: 0.000 min
Response: 720714403
Conc: 733.94 ng/ml



#42 AR-1268-2


R.T.: 7.665 min
Delta R.T.: 0.001 min
Response: 399207129
Conc: 728.24 ng/ml



#43 AR-1268-3


R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 591730207
Conc: 733.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#43 AR-1268-3

R.T.: 7.873 min
Delta R.T.: 0.000 min
Response: 318543602
Conc: 728.44 ng/ml




#44 AR-1268-4

R.T.: 8.196 min
Delta R.T.: 0.000 min
Response: 237470318
Conc: 710.46 ng/ml



#44 AR-1268-4


R.T.: 8.157 min
Delta R.T.: 0.000 min
Response: 125030973
Conc: 719.72 ng/ml



#45 AR-1268-5

R.T.: 8.495 min
Delta R.T.: 0.000 min
Response: 1671425593
Conc: 739.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750



#45 AR-1268-5

R.T.: 8.452 min
Delta R.T.: 0.001 min
Response: 854629551
Conc: 735.80 ng/ml