

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020425\
 Data File : P0109417.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2025 12:12
 Operator : YP/AJ
 Sample : Q1241-07RE
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 JPP-5.3-013025RE

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/05/2025
 Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 01:46:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0020325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 04 08:55:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	3.698	3.694	152.1E6	101.9E6	20.070m	19.072
2) SA Decachlor...	8.757	8.707	70826658	36719061	10.928	11.783

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020425\
Data File : PO109417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2025 12:12
Operator : YP/AJ
Sample : Q1241-07RE
Misc :
ALS Vial : 46 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
JPP-5.3-013025RE


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/05/2025
Supervised By :Ankita Jodhani 02/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 01:46:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO020325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 04 08:55:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.698 min
Delta R.T.: -0.001 min
Response: 152104803
Conc: 20.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
JPP-5.3-013025RE


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/05/2025
Supervised By :Ankita Jodhani 02/05/2025




#1 Tetrachloro-m-xylene

R.T.: 3.694 min
Delta R.T.: 0.000 min
Response: 101858775
Conc: 19.07 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.757 min
Delta R.T.: 0.000 min
Response: 70826658
Conc: 10.93 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.707 min
Delta R.T.: 0.000 min
Response: 36719061
Conc: 11.78 ng/ml