

Surrogate Summary

SDG No.: Q1241

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093725.D	PIBLK-PL093725.D	Decachlorobiphenyl	1	20	22.1	111		43	140
		Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
		Tetrachloro-m-xylene	2	20	20.5	103		77	126
I.BLK-PL093928.D	PIBLK-PL093928.D	Decachlorobiphenyl	1	20	20.6	103		43	140
		Tetrachloro-m-xylene	1	20	21.1	106		77	126
		Decachlorobiphenyl	2	20	20.4	102		43	140
		Tetrachloro-m-xylene	2	20	19.8	99		77	126
PB166413BL	PB166413BL	Decachlorobiphenyl	1	20	24.6	123		10	148
		Tetrachloro-m-xylene	1	20	23.8	119		10	159
		Decachlorobiphenyl	2	20	24.1	121		10	148
		Tetrachloro-m-xylene	2	20	22.1	110		10	159
PB166413BS	PB166413BS	Decachlorobiphenyl	1	20	21.2	106		10	148
		Tetrachloro-m-xylene	1	20	20.0	100		10	159
		Decachlorobiphenyl	2	20	21.7	109		10	148
		Tetrachloro-m-xylene	2	20	18.6	93		10	159
I.BLK-PL093942.D	PIBLK-PL093942.D	Decachlorobiphenyl	1	20	22.6	113		43	140
		Tetrachloro-m-xylene	1	20	21.7	109		77	126
		Decachlorobiphenyl	2	20	22.3	111		43	140
		Tetrachloro-m-xylene	2	20	20.1	101		77	126
I.BLK-PL093957.D	PIBLK-PL093957.D	Decachlorobiphenyl	1	20	20.4	102		43	140
		Tetrachloro-m-xylene	1	20	22.0	110		77	126
		Decachlorobiphenyl	2	20	17.1	86		43	140
		Tetrachloro-m-xylene	2	20	20.8	104		77	126
Q1239-10MS	357MS	Decachlorobiphenyl	1	20	14.9	75		10	148
		Tetrachloro-m-xylene	1	20	19.1	95		10	159
		Decachlorobiphenyl	2	20	13.7	69		10	148
		Tetrachloro-m-xylene	2	20	18.0	90		10	159
Q1239-10MSD	357MSD	Decachlorobiphenyl	1	20	15.4	77		10	148
		Tetrachloro-m-xylene	1	20	18.7	93		10	159
		Decachlorobiphenyl	2	20	13.6	68		10	148
		Tetrachloro-m-xylene	2	20	17.9	90		10	159
Q1241-03	JPP-3.5-013025	Decachlorobiphenyl	1	20	26.7	133		10	148
		Tetrachloro-m-xylene	1	20	11.9	60		10	159
		Decachlorobiphenyl	2	20	24.4	122		10	148
		Tetrachloro-m-xylene	2	20	13.2	66		10	159
Q1241-07	JPP-5.3-013025	Decachlorobiphenyl	1	20	11.2	56		10	148
		Tetrachloro-m-xylene	1	20	14.3	72		10	159
		Decachlorobiphenyl	2	20	9.54	48		10	148
		Tetrachloro-m-xylene	2	20	15.3	76		10	159
Q1241-11	JPP-5.2-013025	Decachlorobiphenyl	1	20	12.8	64		10	148

Surrogate Summary

SDG No.: Q1241

Client: RU2 Engineering, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1241-11	JPP-5.2-013025	Tetrachloro-m-xylene	1	20	13.6	68		10	159
		Decachlorobiphenyl	2	20	11.2	56		10	148
Q1241-15	JPP-5.4-013025	Tetrachloro-m-xylene	2	20	13.9	70		10	159
		Decachlorobiphenyl	1	20	14.4	72		10	148
		Tetrachloro-m-xylene	1	20	14.2	71		10	159
		Decachlorobiphenyl	2	20	11.7	58		10	148
		Tetrachloro-m-xylene	2	20	15.9	79		10	159
Q1241-19	JPP-51.4-013025	Decachlorobiphenyl	1	20	16.9	85		10	148
		Tetrachloro-m-xylene	1	20	16.3	81		10	159
		Decachlorobiphenyl	2	20	14.9	74		10	148
		Tetrachloro-m-xylene	2	20	16.6	83		10	159
I.BLK-PL093970.D	PIBLK-PL093970.D	Decachlorobiphenyl	1	20	19.3	96		43	140
		Tetrachloro-m-xylene	1	20	21.3	107		77	126
		Decachlorobiphenyl	2	20	16.1	81		43	140
		Tetrachloro-m-xylene	2	20	20.0	100		77	126