

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241

Instrument ID: ECD_S

Calibration Date(s): 01/14/2025 01/14/2025

Calibration Times: 10:31 12:07

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028901.D</u>	CF 500 =	<u>PS028902.D</u>		
CF 750 =		<u>PS028903.D</u>	CF 1000 =	<u>PS028904.D</u>	CF 1500 =	<u>PS028905.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	21246200000	19217800000	18444300000	17622300000	16707400000	18647600000	9
2,4-D	3794730000	3389210000	3238030000	3095840000	2967500000	3297060000	10
2,4-DCAA	3179220000	2766210000	2659700000	2530920000	2413760000	2709960000	11

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: RUTW01
Lab Code: CHEM **Case No.:** Q1241 **SAS No.:** Q1241 **SDG NO.:** Q1241
Instrument ID: ECD_S **Calibration Date(s):** 01/14/2025 01/14/2025
Calibration Times: 10:31 12:07

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028901.D</u>	CF 500 =	<u>PS028902.D</u>		
CF 750 =		<u>PS028903.D</u>	CF 1000 =	<u>PS028904.D</u>	CF 1500 =	<u>PS028905.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	9615710000	9419870000	9409010000	9233020000	9015720000	9338670000	2
2,4-D	1602310000	1486700000	1468930000	1440130000	1429250000	1485460000	5
2,4-DCAA	1189550000	1103610000	1095350000	1074740000	1070080000	1106670000	4