

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 02/03/2025

Lab Sample No.: HSTDCCC750 Data File : PS029047.D Time Analyzed: 10:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.180	9.089	9.289	723.740	712.500	1.6
2,4-D	8.309	8.216	8.416	700.140	705.000	-0.7
2,4-DCAA	7.191	7.097	7.297	744.770	750.000	-0.7

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 02/03/2025

Lab Sample No.: HSTDCCC750 Data File : PS029047.D Time Analyzed: 10:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.802	9.713	9.913	708.950	712.500	-0.5
2,4-D	8.903	8.813	9.013	652.250	705.000	-7.5
2,4-DCAA	7.669	7.577	7.777	675.660	750.000	-9.9