

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL094038.D Time Analyzed: 09:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.056	8.953	9.153	47.740	50.000	-4.5
Endrin	6.573	6.472	6.672	48.440	50.000	-3.1
gamma-BHC (Lindane)	4.328	4.227	4.427	51.900	50.000	3.8
Heptachlor	4.916	4.814	5.014	51.240	50.000	2.5
Heptachlor epoxide	5.684	5.582	5.782	50.560	50.000	1.1
Methoxychlor	7.500	7.398	7.598	48.720	50.000	-2.6
Tetrachloro-m-xylene	3.539	3.439	3.639	51.420	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/04/2025

Lab Sample No.: PSTDCCC050 Data File : PL094038.D Time Analyzed: 09:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.910	7.810	8.010	52.230	50.000	4.5
Endrin	5.636	5.536	5.736	53.510	50.000	7.0
gamma-BHC (Lindane)	3.607	3.507	3.707	52.640	50.000	5.3
Heptachlor	3.945	3.845	4.045	50.310	50.000	0.6
Heptachlor epoxide	4.727	4.627	4.827	52.780	50.000	5.6
Methoxychlor	6.609	6.509	6.709	48.950	50.000	-2.1
Tetrachloro-m-xylene	2.774	2.674	2.874	52.080	50.000	4.2