

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG No.: Q1241

Instrument ID: BNA_F Calibration Date/Time: 01/31/2025 17:19

Lab File ID: BF141336.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.303	1.218		-6.5	
Benzaldehyde	0.960	0.972		1.3	
Phenol-d6	1.657	1.534		-7.4	
Phenol	1.757	1.652		-6.0	20.0
bis(2-Chloroethyl)ether	1.347	1.267		-5.9	
2-Chlorophenol	1.396	1.307		-6.4	
2-Methylphenol	1.135	1.062		-6.4	
2,2-oxybis(1-Chloropropane)	2.342	2.199		-6.1	
Acetophenone	0.475	0.424		-10.7	
3+4-Methylphenols	1.393	1.318		-5.4	
n-Nitroso-di-n-propylamine	0.985	0.946	0.050	-4.0	
Nitrobenzene-d5	0.379	0.348		-8.2	
Hexachloroethane	0.571	0.536		-6.1	
Nitrobenzene	0.393	0.365		-7.1	
Isophorone	0.656	0.622		-5.2	
2-Nitrophenol	0.185	0.175		-5.4	20.0
2,4-Dimethylphenol	0.236	0.224		-5.1	
bis(2-Chloroethoxy)methane	0.418	0.386		-7.7	
2,4-Dichlorophenol	0.289	0.272		-5.9	20.0
Naphthalene	1.057	0.966		-8.6	
4-Chloroaniline	0.358	0.329		-8.1	
Hexachlorobutadiene	0.194	0.179		-7.7	20.0
Caprolactam	0.094	0.091		-3.2	
4-Chloro-3-methylphenol	0.310	0.341		10.0	20.0
2-Methylnaphthalene	0.672	0.618		-8.0	
Hexachlorocyclopentadiene	0.168	0.159	0.050	-5.4	
2,4,6-Trichlorophenol	0.372	0.365		-1.9	20.0
2-Fluorobiphenyl	1.299	1.148		-11.6	
2,4,5-Trichlorophenol	0.405	0.381		-5.9	
1,1-Biphenyl	1.560	1.427		-8.5	
2-Chloronaphthalene	1.216	1.102		-9.4	
2-Nitroaniline	0.382	0.366		-4.2	
Dimethylphthalate	1.351	1.290		-4.5	
Acenaphthylene	1.702	1.577		-7.3	
2,6-Dinitrotoluene	0.314	0.300		-4.5	
3-Nitroaniline	0.307	0.292		-4.9	
Acenaphthene	1.216	1.127		-7.3	20.0
2,4-Dinitrophenol	0.146	0.150	0.050	2.7	
4-Nitrophenol	0.225	0.300	0.050	33.3	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG No.: Q1241

Instrument ID: BNA_F Calibration Date/Time: 01/31/2025 17:19

Lab File ID: BF141336.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.627	1.481		-9.0	
2,4-Dinitrotoluene	0.406	0.427		5.2	
Diethylphthalate	1.310	1.249		-4.7	
4-Chlorophenyl-phenylether	0.617	0.562		-8.9	
Fluorene	1.262	1.151		-8.8	
4-Nitroaniline	0.302	0.288		-4.6	
4,6-Dinitro-2-methylphenol	0.118	0.131		11.0	
n-Nitrosodiphenylamine	0.647	0.627		-3.1	20.0
2,4,6-Tribromophenol	0.183	0.177		-3.3	
4-Bromophenyl-phenylether	0.224	0.220		-1.8	
Hexachlorobenzene	0.238	0.232		-2.5	
Atrazine	0.216	0.184		-14.8	
Pentachlorophenol	0.139	0.186		33.8	20.0
Phenanthrene	1.075	1.016		-5.5	
Anthracene	1.053	0.999		-5.1	
Carbazole	0.934	0.892		-4.5	
Di-n-butylphthalate	1.132	1.116		-1.4	
Fluoranthene	1.062	1.011		-4.8	20.0
Pyrene	1.920	1.949		1.5	
Terphenyl-d14	1.297	1.177		-9.3	
Butylbenzylphthalate	0.671	0.656		-2.2	
3,3-Dichlorobenzidine	0.439	0.397		-9.6	
Benzo (a) anthracene	1.380	1.181		-14.4	
Chrysene	1.265	1.179		-6.8	
Bis(2-ethylhexyl)phthalate	0.851	0.815		-4.2	
Di-n-octyl phthalate	1.321	1.285		-2.7	20.0
Benzo (b) fluoranthene	1.258	1.168		-7.2	
Benzo (k) fluoranthene	1.114	1.042		-6.5	
Benzo (a) pyrene	1.063	0.991		-6.8	20.0
Indeno (1,2,3-cd) pyrene	1.291	1.238		-4.1	
Dibenzo (a,h) anthracene	1.039	0.999		-3.8	
Benzo (g,h,i) perylene	1.080	1.050		-2.8	
1,2,4,5-Tetrachlorobenzene	0.569	0.515		-9.5	
1,4-Dioxane	0.573	0.545		-4.9	20.0
2,3,4,6-Tetrachlorophenol	0.319	0.310		-2.8	

All other compounds must meet a minimum RRF of 0.010.