

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG No.: Q1241

Instrument ID: BNA_F Calibration Date/Time: 02/05/2025 10:02

Lab File ID: BF141435.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.303	1.295		-0.6	
Benzaldehyde	0.960	0.988		2.9	
Phenol-d6	1.657	1.600		-3.4	
Phenol	1.757	1.663		-5.3	20.0
bis(2-Chloroethyl)ether	1.347	1.263		-6.2	
2-Chlorophenol	1.396	1.383		-0.9	
2-Methylphenol	1.135	1.083		-4.6	
2,2-oxybis(1-Chloropropane)	2.342	2.038		-13.0	
Acetophenone	0.475	0.493		3.8	
3+4-Methylphenols	1.393	1.395		0.1	
n-Nitroso-di-n-propylamine	0.985	0.951	0.050	-3.5	
Nitrobenzene-d5	0.379	0.397		4.5	
Hexachloroethane	0.571	0.582		1.9	
Nitrobenzene	0.393	0.403		2.5	
Isophorone	0.656	0.632		-3.7	
2-Nitrophenol	0.185	0.191		3.2	20.0
2,4-Dimethylphenol	0.236	0.229		-3.0	
bis(2-Chloroethoxy)methane	0.418	0.393		-6.0	
2,4-Dichlorophenol	0.289	0.297		2.8	20.0
Naphthalene	1.057	1.058		0.1	
4-Chloroaniline	0.358	0.335		-6.4	
Hexachlorobutadiene	0.194	0.212		9.3	20.0
Caprolactam	0.094	0.092		-2.1	
4-Chloro-3-methylphenol	0.310	0.323		4.2	20.0
2-Methylnaphthalene	0.672	0.670		-0.3	
Hexachlorocyclopentadiene	0.168	0.166	0.050	-1.2	
2,4,6-Trichlorophenol	0.372	0.385		3.5	20.0
2-Fluorobiphenyl	1.299	1.338		3.0	
2,4,5-Trichlorophenol	0.405	0.418		3.2	
1,1-Biphenyl	1.560	1.536		-1.5	
2-Chloronaphthalene	1.216	1.222		0.5	
2-Nitroaniline	0.382	0.396		3.7	
Dimethylphthalate	1.351	1.345		-0.4	
Acenaphthylene	1.702	1.696		-0.4	
2,6-Dinitrotoluene	0.314	0.319		1.6	
3-Nitroaniline	0.307	0.309		0.7	
Acenaphthene	1.216	1.232		1.3	20.0
2,4-Dinitrophenol	0.146	0.160	0.050	9.6	
4-Nitrophenol	0.225	0.235	0.050	4.4	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG No.: Q1241

Instrument ID: BNA_F Calibration Date/Time: 02/05/2025 10:02

Lab File ID: BF141435.D Init. Calib. Date(s): 01/29/2025 01/29/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.627	1.626		-0.1	
2,4-Dinitrotoluene	0.406	0.428		5.4	
Diethylphthalate	1.310	1.326		1.2	
4-Chlorophenyl-phenylether	0.617	0.644		4.4	
Fluorene	1.262	1.312		4.0	
4-Nitroaniline	0.302	0.314		4.0	
4,6-Dinitro-2-methylphenol	0.118	0.135		14.4	
n-Nitrosodiphenylamine	0.647	0.657		1.5	20.0
2,4,6-Tribromophenol	0.183	0.197		7.7	
4-Bromophenyl-phenylether	0.224	0.231		3.1	
Hexachlorobenzene	0.238	0.246		3.4	
Atrazine	0.216	0.200		-7.4	
Pentachlorophenol	0.139	0.151		8.6	20.0
Phenanthrene	1.075	1.115		3.7	
Anthracene	1.053	1.101		4.6	
Carbazole	0.934	1.005		7.6	
Di-n-butylphthalate	1.132	1.165		2.9	
Fluoranthene	1.062	1.154		8.7	20.0
Pyrene	1.920	1.874		-2.4	
Terphenyl-d14	1.297	1.290		-0.5	
Butylbenzylphthalate	0.671	0.667		-0.6	
3,3-Dichlorobenzidine	0.439	0.419		-4.6	
Benzo (a) anthracene	1.380	1.310		-5.1	
Chrysene	1.265	1.242		-1.8	
Bis(2-ethylhexyl)phthalate	0.851	0.821		-3.5	
Di-n-octyl phthalate	1.321	1.243		-5.9	20.0
Benzo (b) fluoranthene	1.258	1.327		5.5	
Benzo (k) fluoranthene	1.114	1.082		-2.9	
Benzo (a) pyrene	1.063	1.102		3.7	20.0
Indeno (1,2,3-cd) pyrene	1.291	1.411		9.3	
Dibenzo (a,h) anthracene	1.039	1.133		9.0	
Benzo (g,h,i) perylene	1.080	1.191		10.3	
1,2,4,5-Tetrachlorobenzene	0.569	0.579		1.8	
1,4-Dioxane	0.573	0.500		-12.7	20.0
2,3,4,6-Tetrachlorophenol	0.319	0.335		5.0	

All other compounds must meet a minimum RRF of 0.010.