

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG No.: Q1241

Instrument ID: BNA_F Calibration Date/Time: 02/07/2025 14:33

Lab File ID: BF141506.D Init. Calib. Date(s): 02/06/2025 02/06/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:07 14:14

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.276	1.294		1.4	
Benzaldehyde	0.857	0.844		-1.5	
Phenol-d6	1.612	1.599		-0.8	
Phenol	1.678	1.676		-0.1	20.0
bis(2-Chloroethyl)ether	1.261	1.237		-1.9	
2-Chlorophenol	1.378	1.383		0.4	
2-Methylphenol	1.079	1.064		-1.4	
2,2-oxybis(1-Chloropropane)	2.158	2.121		-1.7	
Acetophenone	0.498	0.489		-1.8	
3+4-Methylphenols	1.371	1.329		-3.1	
n-Nitroso-di-n-propylamine	0.964	0.978	0.050	1.5	
Nitrobenzene-d5	0.383	0.384		0.3	
Hexachloroethane	0.550	0.545		-0.9	
Nitrobenzene	0.374	0.372		-0.5	
Isophorone	0.611	0.607		-0.7	
2-Nitrophenol	0.186	0.191		2.7	20.0
2,4-Dimethylphenol	0.217	0.219		0.9	
bis(2-Chloroethoxy)methane	0.392	0.396		1.0	
2,4-Dichlorophenol	0.294	0.296		0.7	20.0
Naphthalene	1.041	1.040		-0.1	
4-Chloroaniline	0.363	0.327		-9.9	
Hexachlorobutadiene	0.192	0.191		-0.5	20.0
Caprolactam	0.089	0.093		4.5	
4-Chloro-3-methylphenol	0.325	0.326		0.3	20.0
2-Methylnaphthalene	0.677	0.668		-1.3	
Hexachlorocyclopentadiene	0.192	0.193	0.050	0.5	
2,4,6-Trichlorophenol	0.374	0.368		-1.6	20.0
2-Fluorobiphenyl	1.298	1.268		-2.3	
2,4,5-Trichlorophenol	0.405	0.399		-1.5	
1,1-Biphenyl	1.551	1.528		-1.5	
2-Chloronaphthalene	1.147	1.132		-1.3	
2-Nitroaniline	0.385	0.385		0.0	
Dimethylphthalate	1.358	1.338		-1.5	
Acenaphthylene	1.705	1.697		-0.5	
2,6-Dinitrotoluene	0.297	0.295		-0.7	
3-Nitroaniline	0.319	0.315		-1.3	
Acenaphthene	1.147	1.116		-2.7	20.0
2,4-Dinitrophenol	0.176	0.169	0.050	-4.0	
4-Nitrophenol	0.237	0.240	0.050	1.3	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG No.: Q1241

Instrument ID: BNA_F Calibration Date/Time: 02/07/2025 14:33

Lab File ID: BF141506.D Init. Calib. Date(s): 02/06/2025 02/06/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:07 14:14

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.683	1.655		-1.7	
2,4-Dinitrotoluene	0.395	0.400		1.3	
Diethylphthalate	1.336	1.313		-1.7	
4-Chlorophenyl-phenylether	0.626	0.616		-1.6	
Fluorene	1.301	1.278		-1.8	
4-Nitroaniline	0.324	0.326		0.6	
4,6-Dinitro-2-methylphenol	0.139	0.139		0.0	
n-Nitrosodiphenylamine	0.640	0.631		-1.4	20.0
2,4,6-Tribromophenol	0.201	0.200		-0.5	
4-Bromophenyl-phenylether	0.224	0.217		-3.1	
Hexachlorobenzene	0.230	0.229		-0.4	
Atrazine	0.192	0.182		-5.2	
Pentachlorophenol	0.147	0.147		0.0	20.0
Phenanthrene	1.101	1.082		-1.7	
Anthracene	1.107	1.095		-1.1	
Carbazole	0.996	1.004		0.8	
Di-n-butylphthalate	1.150	1.152		0.2	
Fluoranthene	1.167	1.168		0.1	20.0
Pyrene	1.703	1.702		-0.1	
Terphenyl-d14	1.185	1.171		-1.2	
Butylbenzylphthalate	0.590	0.640		8.5	
3,3-Dichlorobenzidine	0.381	0.385		1.0	
Benzo (a) anthracene	1.329	1.316		-1.0	
Chrysene	1.228	1.167		-5.0	
Bis(2-ethylhexyl)phthalate	0.665	0.759		14.1	
Di-n-octyl phthalate	0.949	1.017		7.2	20.0
Benzo (b) fluoranthene	1.343	1.426		6.2	
Benzo (k) fluoranthene	1.147	1.070		-6.7	
Benzo (a) pyrene	1.087	1.090		0.3	20.0
Indeno (1,2,3-cd) pyrene	1.236	1.245		0.7	
Dibenzo (a,h) anthracene	1.001	1.011		1.0	
Benzo (g,h,i) perylene	1.048	1.028		-1.9	
1,2,4,5-Tetrachlorobenzene	0.579	0.571		-1.4	
1,4-Dioxane	0.513	0.512		-0.2	20.0
2,3,4,6-Tetrachlorophenol	0.349	0.346		-0.9	

All other compounds must meet a minimum RRF of 0.010.