

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/10/2025

Lab File ID: BF141531.D Time Analyzed: 11:34

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	91504	6.792	356056	8.08	192109	9.83
UPPER LIMIT	183008	7.292	712112	8.575	384218	10.327
LOWER LIMIT	45752	6.292	178028	7.575	96054.5	9.327
EPA SAMPLE NO.						
01 JPP-3.5-013025DL	76076	6.79	302314	8.07	158827	9.82
02 JPP-5.4-013025DL	77702	6.79	310551	8.07	156153	9.82

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.