

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM

SAS No.: Q1241 SDG NO.: Q1241

Lab File ID: BF141434.D

DFTPP Injection Date: 02/05/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:36

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	42.7
68	Less than 2.0% of mass 69	0.7 (2) 1
69	Mass 69 relative abundance	35.1
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	48.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.4
275	10.0 - 60.0% of mass 198	26.8
365	Greater than 1% of mass 198	2.8
441	Present, but less than mass 443	12.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.5 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF141435.D	02/05/2025	10:02
PB166418BL	PB166418BL	BF141440.D	02/05/2025	12:13
PB166418BS	PB166418BS	BF141441.D	02/05/2025	12:39