

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: RUTW01

Lab Code: CHEM

SAS No.: Q1241 SDG NO.: Q1241

Lab File ID: BF141530.D

DFTPP Injection Date: 02/10/2025

Instrument ID: BNA_F

DFTPP Injection Time: 11:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	43.6
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	37.3
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	49.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	27.1
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	12.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	15.4 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF141531.D	02/10/2025	11:34
JPP-3.5-013025DL	Q1241-03DL	BF141546.D	02/10/2025	20:19
JPP-5.4-013025DL	Q1241-15DL	BF141549.D	02/10/2025	21:38