

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG No.: Q1241
 Instrument ID: BNA_F Calibration Date/Time: 02/04/2025 10:42
 Lab File ID: BF141398.D Init. Calib. Date(s): 01/29/2025 01/29/2025
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 17:05
 GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.380	1.270		-8.0	
2-Fluorophenol	1.303	1.302		-0.1	
Phenol-d6	1.657	1.560		-5.9	
1,4-Dichlorobenzene	1.547	1.539		-0.5	20.0
2-Methylphenol	1.135	1.074		-5.4	
3+4-Methylphenols	1.393	1.375		-1.3	
Nitrobenzene-d5	0.379	0.391		3.2	
Hexachloroethane	0.571	0.583		2.1	
Nitrobenzene	0.393	0.398		1.3	
Hexachlorobutadiene	0.194	0.208		7.2	20.0
2,4,6-Trichlorophenol	0.372	0.382		2.7	20.0
2-Fluorobiphenyl	1.299	1.328		2.2	
2,4,5-Trichlorophenol	0.405	0.421		4.0	
2,4-Dinitrotoluene	0.406	0.441		8.6	
2,4,6-Tribromophenol	0.183	0.202		10.4	
Hexachlorobenzene	0.238	0.243		2.1	
Pentachlorophenol	0.139	0.161		15.8	20.0
Terphenyl-d14	1.297	1.200		-7.5	

All other compounds must meet a minimum RRF of 0.010.