

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1241

Client: RU2 Engineering, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1239-02MSD	Client Sample ID:	286MSD					DataFile:	BF141414.D		
Pyridine	500	0	330	ug/L	66		10		10	109	20
1,4-Dichlorobenzene	500	0	480	ug/L	96		6		55	125	20
2-Methylphenol	500	0	480	ug/L	96		6		37	126	20
3+4-Methylphenols	500	0	500	ug/L	100		6		31	127	20
Hexachloroethane	500	0	500	ug/L	100		6		49	110	20
Nitrobenzene	500	0	470	ug/L	94		2		62	112	20
Hexachlorobutadiene	500	0	500	ug/L	100		2		52	125	20
2,4,6-Trichlorophenol	500	0	500	ug/L	100		4		78	112	20
2,4,5-Trichlorophenol	500	0	510	ug/L	102		4		71	111	20
2,4-Dinitrotoluene	500	0	540	ug/L	108		6		50	142	20
Hexachlorobenzene	500	0	500	ug/L	100		4		72	115	20
Pentachlorophenol	1000	0	1200	ug/L	120		9		25	139	20