

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1241

Client: RU2 Engineering, LLC

Analytical Method: 8270E

DataFile: BF141400.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB166483BS	Pyridine	50	39.3	ug/L	79				29	97	
	1,4-Dichlorobenzene	50	44.7	ug/L	89				76	103	
	2-Methylphenol	50	43.6	ug/L	87				69	109	
	3+4-Methylphenols	50	46.4	ug/L	93				67	106	
	Hexachloroethane	50	45.0	ug/L	90				76	118	
	Nitrobenzene	50	43.8	ug/L	88				58	106	
	Hexachlorobutadiene	50	46.6	ug/L	93				69	101	
	2,4,6-Trichlorophenol	50	47.5	ug/L	95				61	110	
	2,4,5-Trichlorophenol	50	45.8	ug/L	92				70	106	
	2,4-Dinitrotoluene	50	47.5	ug/L	95				60	115	
	Hexachlorobenzene	50	47.0	ug/L	94				73	106	
	Pentachlorophenol	100	110	ug/L	110				47	114	