

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241
 Lab File ID: VN085617.D Date Analyzed: 02/03/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:56
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	193909	8.22	331362	9.10	301323	11.87
UPPER LIMIT	387818	8.718	662724	9.6	602646	12.365
LOWER LIMIT	96954.5	7.718	165681	8.6	150662	11.365
EPA SAMPLE NO.						
JPP-5.3-013025	164691	8.22	346323	9.10	326963	11.87
JPP-5.2-013025	157669	8.22	330403	9.10	315638	11.87
JPP-5.4-013025	163991	8.22	343832	9.10	330622	11.87
JPP-51.4-013025	168419	8.22	349686	9.10	332324	11.87
VN0203WBL01	183958	8.22	354826	9.10	307130	11.87
VN0203WBS01	188870	8.22	327054	9.10	287245	11.87
VN0203WBSD01	177381	8.22	316818	9.10	277592	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.