

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241
 Lab File ID: VN085645.D Date Analyzed: 02/04/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:05
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	200704	8.22	345148	9.10	309053	11.87
UPPER LIMIT	401408	8.724	690296	9.6	618106	12.365
LOWER LIMIT	100352	7.724	172574	8.6	154527	11.365
EPA SAMPLE NO.						
JPP-3.5-013025	192282	8.22	373073	9.10	340715	11.87
VN0204WBL01	197983	8.22	379434	9.10	328690	11.87
VN0204WBS02	206401	8.22	367320	9.10	317764	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.