

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY020325

Review By	Mahesh Dadoda	Review On	2/4/2025 3:47:24 PM				
Supervise By	Semsettin Yesilyurt	Supervise On	2/4/2025 3:53:22 PM				
SubDirectory	VY020325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y020325s.m		
STD. NAME		STD REF.#					
Tune/Reschk		VP132839					
Initial Calibration Stds		VP132837,VP132841,VP132843,VP132845,VP132847,VP132849					
CCC		VP132853					
Internal Standard/PEM		VP131783					
ICV/I.BLK		VP132851					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021019.D	03 Feb 2025 09:51		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021020.D	03 Feb 2025 10:35	Good for DOD	SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY021021.D	03 Feb 2025 10:57		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021022.D	03 Feb 2025 11:20		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021023.D	03 Feb 2025 11:43		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021024.D	03 Feb 2025 12:21		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021025.D	03 Feb 2025 12:44		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021026.D	03 Feb 2025 13:11		SY/MD	Ok
9	VSTDICV050	ICVVY020325	VY021027.D	03 Feb 2025 13:37		SY/MD	Ok,M
10	VY0203SBL01	VY0203SBL01	VY021028.D	03 Feb 2025 14:13		SY/MD	Ok
11	VY0203SBS01	VY0203SBS01	VY021029.D	03 Feb 2025 14:55		SY/MD	Ok,M
12	VY0203SBSD01	VY0203SBSD01	VY021030.D	03 Feb 2025 15:17		SY/MD	Ok,M
13	Q1239-10	357	VY021031.D	03 Feb 2025 15:50	vial B Not purged	SY/MD	Not Ok
14	Q1215-01	JPP-29.1-012825	VY021032.D	03 Feb 2025 16:14	vial A ISTD Fail	SY/MD	Ok
15	Q1215-05	JPP-29.2-012825	VY021033.D	03 Feb 2025 16:37	Not purged vial-A	SY/MD	Not Ok
16	Q1216-01	JPP-18.1-012825	VY021034.D	03 Feb 2025 17:01	vial A Internal Standard Fail	SY/MD	ReRun
17	Q1216-05	JPP-21.1-012825	VY021035.D	03 Feb 2025 17:24	vial B ISTD Fail	SY/MD	Ok
18	IBLK	IBLK	VY021036.D	03 Feb 2025 17:47	Not purged	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY020325

Review By	Mahesh Dadoda	Review On	2/4/2025 3:47:24 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/4/2025 3:53:22 PM		
SubDirectory	VY020325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y020325s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP132839			
Initial Calibration Stds		VP132837,VP132841,VP132843,VP132845,VP132847,VP132849			
CCC		VP132853			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP132851			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	IBLK	IBLK	VY021037.D	03 Feb 2025 18:11	Not purged	SY/MD	Not Ok
20	Q1216-17	JPP-26.2-012825	VY021038.D	03 Feb 2025 18:34	vial B Internal Standard Fail	SY/MD	Ok
21	Q1206-01	JPP-20.1-012725	VY021039.D	03 Feb 2025 18:58	vial B	SY/MD	Ok
22	Q1206-05	JPP-16.3-012725	VY021040.D	03 Feb 2025 19:21	vial B	SY/MD	Ok
23	Q1207-01	JPP-2.1-012725	VY021041.D	03 Feb 2025 19:44	vial-B	SY/MD	Ok
24	Q1207-05	JPP-5.1-012725	VY021042.D	03 Feb 2025 20:08	vial-B	SY/MD	Ok
25	Q1207-09	JPP-4.5-012725	VY021043.D	03 Feb 2025 20:31	vial-B	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY021044.D	03 Feb 2025 20:54		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY020425

Review By	Maresh Dadoda	Review On	2/6/2025 11:05:18 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/6/2025 11:07:34 AM
SubDirectory	VY020425	HP Acquire Method	HP Processing Method 82y020325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132861
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132865,VP132866 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021045.D	04 Feb 2025 09:16		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021046.D	04 Feb 2025 09:47		SY/MD	Ok,M
3	VY0204SBL01	VY0204SBL01	VY021047.D	04 Feb 2025 10:15		SY/MD	Ok
4	VY0204SBS01	VY0204SBS01	VY021048.D	04 Feb 2025 10:52		SY/MD	Ok,M
5	VY0204SBSD01	VY0204SBSD01	VY021049.D	04 Feb 2025 11:15		SY/MD	Ok,M
6	Q1243-01	CL-01-01302025	VY021050.D	04 Feb 2025 11:55	vial-A	SY/MD	Ok
7	Q1207-17	JPP-20.2-012725	VY021051.D	04 Feb 2025 12:19	vial-B	SY/MD	Ok
8	Q1207-13	JPP-16.2-012725	VY021052.D	04 Feb 2025 12:42	vial-B	SY/MD	Ok
9	Q1216-01	JPP-18.1-012825	VY021053.D	04 Feb 2025 13:06	vial-B	SY/MD	Ok
10	Q1215-01	JPP-29.1-012825	VY021054.D	04 Feb 2025 13:29	NOT PURGE vial-B	SY/MD	Not Ok
11	Q1215-05	JPP-29.2-012825	VY021055.D	04 Feb 2025 13:52	Internal Standard Fail vial-B	SY/MD	Ok
12	Q1232-01	JPP-46.2-012925	VY021056.D	04 Feb 2025 14:16	vial-A	SY/MD	Ok
13	Q1232-05	JPP-46.1-012925	VY021057.D	04 Feb 2025 14:39	vial-A	SY/MD	Ok
14	Q1232-09	JPP-42.1-012925	VY021058.D	04 Feb 2025 15:03	vial-A	SY/MD	Ok
15	Q1232-13	JPP-42.2-012925	VY021059.D	04 Feb 2025 15:26	vial-A	SY/MD	Ok
16	Q1232-17	JPP-51.1-012925	VY021060.D	04 Feb 2025 15:50	vial-A	SY/MD	Ok
17	Q1235-01	JPP-51.2-012925	VY021061.D	04 Feb 2025 16:13	vial-A	SY/MD	Ok
18	Q1235-05	JPP-16.1-012925	VY021062.D	04 Feb 2025 16:36	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY020425

Review By	Maresh Dadoda	Review On	2/6/2025 11:05:18 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/6/2025 11:07:34 AM
SubDirectory	VY020425	HP Acquire Method	HP Processing Method 82y020325s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132861		
CCC	VP132865,VP132866		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1241-01	JPP-3.5-013025	VY021063.D	04 Feb 2025 17:00	vial-A	SY/MD	Ok
20	Q1241-05	JPP-5.3-013025	VY021064.D	04 Feb 2025 17:23	vial-A Internal Standard Fail	SY/MD	ReRun
21	Q1241-09	JPP-5.2-013025	VY021065.D	04 Feb 2025 17:47	vial-A	SY/MD	Ok
22	Q1241-13	JPP-5.4-013025	VY021066.D	04 Feb 2025 18:10	vial-A	SY/MD	Ok
23	Q1241-17	JPP-5.1.4-013025	VY021067.D	04 Feb 2025 18:33	vial-A	SY/MD	Ok
24	Q1242-01	JPP-6.2-013025	VY021068.D	04 Feb 2025 18:57	vial-A	SY/MD	Ok
25	Q1271-01	RBR200030	VY021069.D	04 Feb 2025 19:20	vial-B	SY/MD	Ok
26	Q1216-09	JPP-21.2-012825	VY021070.D	04 Feb 2025 19:44	vial-B	SY/MD	Ok
27	Q1216-13	JPP-26.1-012825	VY021071.D	04 Feb 2025 20:07	vial-B	SY/MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VY021072.D	04 Feb 2025 20:30		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY020525

Review By	Maresh Dadoda	Review On	2/6/2025 11:06:45 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/6/2025 11:07:48 AM
SubDirectory	VY020525	HP Acquire Method	HP Processing Method 82y020325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP132891
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP132892,VP132893 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021073.D	05 Feb 2025 08:16		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021074.D	05 Feb 2025 09:16		SY/MD	Ok,M
3	VY0205SBL01	VY0205SBL01	VY021075.D	05 Feb 2025 09:51		SY/MD	Ok
4	VY0205SBS01	VY0205SBS01	VY021076.D	05 Feb 2025 10:21		SY/MD	Ok,M
5	VY0205SBSD01	VY0205SBSD01	VY021077.D	05 Feb 2025 10:43		SY/MD	Ok,M
6	Q1284-02	SU-03-2-4-2025	VY021078.D	05 Feb 2025 11:07		SY/MD	Ok
7	Q1269-01	VNJ-231	VY021079.D	05 Feb 2025 11:30	Internal Standard Fail vial-A	SY/MD	ReRun
8	Q1262-01	ETGI-371	VY021080.D	05 Feb 2025 11:54	Internal Standard Fail vial-A	SY/MD	ReRun
9	Q1261-01	CHRT-20430	VY021081.D	05 Feb 2025 12:17	Internal Standard Fail vial-A	SY/MD	ReRun
10	Q1262-03	CONCRETE-PAD	VY021082.D	05 Feb 2025 12:40	vial-A	SY/MD	Ok
11	Q1261-01RE	CHRT-20430RE	VY021083.D	05 Feb 2025 13:04	Internal Standard Fail vial-B	SY/MD	Confirms
12	Q1269-01RE	VNJ-231RE	VY021084.D	05 Feb 2025 13:27	Surrogate Fail;Internal Standard Fail vial-B	SY/MD	Confirms
13	Q1280-01	72-11984	VY021085.D	05 Feb 2025 13:51	Internal Standard Fail vial-B	SY/MD	ReRun
14	Q1277-01	RT 3249-VOC	VY021086.D	05 Feb 2025 14:14	Internal Standard Fail vial-B	SY/MD	Ok
15	Q1289-02	FL-DRUMS-A	VY021087.D	05 Feb 2025 14:38	Surrogate Fail;Internal Standard Fail vial-A	SY/MD	ReRun
16	Q1289-04	FL-DRUMS-B	VY021088.D	05 Feb 2025 15:01	Internal Standard Fail vial-A	SY/MD	ReRun
17	Q1262-01RE	ETGI-371RE	VY021089.D	05 Feb 2025 15:24	Internal Standard Fail vial-B	SY/MD	Confirms

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY020525

Review By	Mahesh Dadoda	Review On	2/6/2025 11:06:45 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/6/2025 11:07:48 AM
SubDirectory	VY020525	HP Acquire Method	HP Processing Method 82y020325s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP132891		
CCC	VP132892,VP132893		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q1288-02	HR-01-020425	VY021090.D	05 Feb 2025 15:48	Internal Standard Fail vial-A	SY/MD	ReRun
19	Q1288-05	HR-04-020425	VY021091.D	05 Feb 2025 16:11	Internal Standard Fail vial-A	SY/MD	ReRun
20	Q1241-05	JPP-5.3-013025	VY021092.D	05 Feb 2025 16:35	vial-B	SY/MD	Ok
21	Q1286-02	NB-07-020425-VOC	VY021093.D	05 Feb 2025 16:58	vial-A	SY/MD	Ok
22	Q1276-02	TR-05-020325	VY021094.D	05 Feb 2025 17:21	vial-A	SY/MD	Ok
23	Q1281-01	OR-02--2-03-2025	VY021095.D	05 Feb 2025 17:45	vial-A	SY/MD	Ok,M
24	Q1285-01	72-11977	VY021096.D	05 Feb 2025 18:08	Internal Standard Fail vial-A	SY/MD	ReRun
25	VIBLK	VIBLK	VY021097.D	05 Feb 2025 18:32	ISTD Fail	SY/MD	Ok
26	VIBLK	VIBLK	VY021098.D	05 Feb 2025 18:55	ISTD Fail	SY/MD	Ok
27	Q1290-02	VOC	VY021099.D	05 Feb 2025 19:18	Not purged well;Surrogate Fail;Internal Standard Fail vial -A	SY/MD	Not Ok
28	VSTDCCC050	VSTDCCC050EC	VY021100.D	05 Feb 2025 19:40		SY/MD	Ok,M

M : Manual Integration