

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RUTW01
 Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241
 Lab File ID: VY021074.D Date Analyzed: 02/05/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:16
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	88650	13.358				
UPPER LIMIT	177300	13.858				
LOWER LIMIT	44325	12.858				
EPA SAMPLE NO.						
JPP-5.3-013025	91291	13.35				
VY0205SBL01	107232	13.36				
VY0205SBS01	102238	13.36				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.