

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: RUTW01  
Lab Code: CHEM Case No.: Q1241 SAS No.: Q1241 SDG NO.: Q1241  
Lab File ID: VY021019.D BFB Injection Date: 02/03/2025  
Instrument ID: MSVOA\_Y BFB Injection Time: 09:51  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 15.3                 |
| 75  | 30.0 - 60.0% of mass 95            | 49                   |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.6                  |
| 173 | Less than 2.0% of mass 174         | 0.3 ( 0.4 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 89.7                 |
| 175 | 5.0 - 9.0% of mass 174             | 7 ( 7.8 ) 1          |
| 176 | 95.0 - 101.0% of mass 174          | 87 ( 97 ) 1          |
| 177 | 5.0 - 9.0% of mass 176             | 5.9 ( 6.8 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDIC005         | VSTDIC005        | VY021020.D     | 02/03/2025       | 10:35            |
| VSTDIC010         | VSTDIC010        | VY021021.D     | 02/03/2025       | 10:57            |
| VSTDIC020         | VSTDIC020        | VY021022.D     | 02/03/2025       | 11:20            |
| VSTDIC050         | VSTDIC050        | VY021023.D     | 02/03/2025       | 11:43            |
| VSTDIC100         | VSTDIC100        | VY021024.D     | 02/03/2025       | 12:21            |
| VSTDIC150         | VSTDIC150        | VY021025.D     | 02/03/2025       | 12:44            |