

Initial and Continuing Calibration Verification

Client: RU2 Engineering, LLC

SDG No.: Q1242

Project: NYCDDC SANTWOBR Brooklyn Bridge BBMCR

RunNo.: LB134515

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.02	7	100	90-110	01/31/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	01/31/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	01/31/2025

Initial and Continuing Calibration Verification

Client: RU2 Engineering, LLC

SDG No.: Q1242

Project: NYCDDC SANTWOBR Brooklyn Bridge BBMCR

RunNo.: LB134524

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.097	0.099	98	85-115	02/03/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	02/03/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	02/03/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	02/03/2025