



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: RU2 Engineering, LLC SDG No.: Q1242
Contract: RUTW01 Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.22	0.2	111	40 - 160	CV	02/04/2025	11:38	LB134551
CRI01	Arsenic	20.6	20.0	103	40 - 160	P	02/06/2025	13:12	Lb134604
	Barium	89.7	100	90	40 - 160	P	02/06/2025	13:12	Lb134604
	Cadmium	5.91	6.0	98	40 - 160	P	02/06/2025	13:12	Lb134604
	Chromium	9.92	10.0	99	40 - 160	P	02/06/2025	13:12	Lb134604
	Lead	11.4	12.0	95	40 - 160	P	02/06/2025	13:12	Lb134604
	Selenium	19.9	20.0	100	40 - 160	P	02/06/2025	13:12	Lb134604
	Silver	10.7	10.0	107	40 - 160	P	02/06/2025	13:12	Lb134604
CRI01	Arsenic	20.3	20.0	102	40 - 160	P	02/07/2025	12:35	LB134636
	Barium	92.8	100	93	40 - 160	P	02/07/2025	12:35	LB134636
	Cadmium	5.87	6.0	98	40 - 160	P	02/07/2025	12:35	LB134636
	Chromium	9.48	10.0	95	40 - 160	P	02/07/2025	12:35	LB134636
	Lead	10.9	12.0	91	40 - 160	P	02/07/2025	12:35	LB134636
	Selenium	19.8	20.0	99	40 - 160	P	02/07/2025	12:35	LB134636
	Silver	10.1	10.0	101	40 - 160	P	02/07/2025	12:35	LB134636