

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG NO.: Q1242

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 09:34 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1242-2 (2)	4.81	4.82	4.72	4.92	0.01
Aroclor-1242-3 (3)	4.87	4.87	4.77	4.97	0.00
Aroclor-1242-4 (4)	4.99	4.99	4.89	5.09	0.00
Aroclor-1242-5 (5)	5.64	5.65	5.55	5.75	0.01
Tetrachloro-m-xylene	3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl	8.76	8.76	8.66	8.86	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG NO.: Q1242

Continuing Calib Date: 01/31/2025 Initial Calibration Date(s): 01/21/2025 01/22/2025

Continuing Calib Time: 09:34 Initial Calibration Time(s): 17:36 01:50

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1242-2 (2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1242-3 (3)	4.97	4.98	4.88	5.08	0.01
Aroclor-1242-4 (4)	5.06	5.06	4.96	5.16	0.00
Aroclor-1242-5 (5)	5.58	5.58	5.48	5.68	0.00
Tetrachloro-m-xylene	3.70	3.70	3.60	3.80	0.01
Decachlorobiphenyl	8.71	8.71	8.61	8.81	0.00