

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG NO.: Q1242

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO109364.D Time Analyzed: 06:43

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.793	4.695	4.895	524.500	500.000	4.9
Aroclor-1016-2	4.812	4.714	4.914	530.300	500.000	6.1
Aroclor-1016-3	4.868	4.770	4.970	516.520	500.000	3.3
Aroclor-1016-4	4.989	4.891	5.091	516.180	500.000	3.2
Aroclor-1016-5	5.247	5.149	5.349	540.580	500.000	8.1
Aroclor-1260-1	6.289	6.190	6.390	422.370	500.000	-15.5
Aroclor-1260-2	6.478	6.379	6.579	408.310	500.000	-18.3
Aroclor-1260-3	6.847	6.747	6.947	379.480	500.000	-24.1
Aroclor-1260-4	7.107	7.008	7.208	388.000	500.000	-22.4
Aroclor-1260-5	7.348	7.249	7.449	382.060	500.000	-23.6
Decachlorobiphenyl	8.757	8.658	8.858	33.880	50.000	-32.2
Tetrachloro-m-xylene	3.698	3.600	3.800	53.260	50.000	6.5

CALIBRATION VERIFICATION SUMMARY

Contract: RUTW01

Lab Code: CHEM Case No.: Q1242 SAS No.: Q1242 SDG NO.: Q1242

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO109364.D Time Analyzed: 06:43

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.778	4.703	4.903	566.230	500.000	13.2
Aroclor-1016-2	4.798	4.722	4.922	529.800	500.000	6.0
Aroclor-1016-3	4.973	4.898	5.098	549.850	500.000	10.0
Aroclor-1016-4	5.015	4.940	5.140	471.870	500.000	-5.6
Aroclor-1016-5	5.229	5.153	5.353	533.700	500.000	6.7
Aroclor-1260-1	6.262	6.186	6.386	485.550	500.000	-2.9
Aroclor-1260-2	6.450	6.373	6.573	458.810	500.000	-8.2
Aroclor-1260-3	6.604	6.527	6.727	431.070	500.000	-13.8
Aroclor-1260-4	7.075	6.999	7.199	405.070	500.000	-19.0
Aroclor-1260-5	7.316	7.239	7.439	399.550	500.000	-20.1
Decachlorobiphenyl	8.709	8.633	8.833	37.230	50.000	-25.5
Tetrachloro-m-xylene	3.695	3.618	3.818	55.990	50.000	12.0