

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1242

Client: RU2 Engineering, LLC

Analytical Method: 8081B

DataFile : PL093961.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: Q1239-10MSD	357MSD											
	alpha-BHC	19.65	0	19.8	ug/kg	101		1		60	144	20
	beta-BHC	19.65	0	20.1	ug/kg	102		1		54	143	20
	delta-BHC	19.65	0	20.0	ug/kg	102		5		47	144	20
	gamma-BHC (Lindane)	19.65	0	19.4	ug/kg	99		1		61	140	20
	Heptachlor	19.65	0	20.3	ug/kg	103		6		63	135	20
	Aldrin	19.65	0	19.5	ug/kg	99		1		49	139	20
	Heptachlor epoxide	19.65	0	20.0	ug/kg	102		0		32	180	20
	Endosulfan I	19.65	0	19.5	ug/kg	99		2		56	142	20
	Dieldrin	19.65	0	20.6	ug/kg	105		1		47	161	20
	4,4'-DDE	19.65	0	21.4	ug/kg	109		0		55	136	20
	Endrin	19.65	0	20.9	ug/kg	106		3		57	139	20
	Endosulfan II	19.65	0	20.1	ug/kg	102		3		40	163	20
	4,4'-DDD	19.65	0.28	22.2	ug/kg	112		1		37	192	20
	Endosulfan sulfate	19.65	0	19.7	ug/kg	100		6		62	139	20
	4,4'-DDT	19.65	0	20.5	ug/kg	104		8		51	146	20
	Methoxychlor	19.65	0	19.8	ug/kg	101		13		54	136	20
	Endrin ketone	19.65	0	18.8	ug/kg	96		8		60	129	20
	Endrin aldehyde	19.65	0	18.7	ug/kg	95		30	*	59	132	20
	alpha-Chlordane	19.65	0	20.6	ug/kg	105		1		30	192	20
	gamma-Chlordane	19.65	0	20.9	ug/kg	106		1		44	175	20